



THE AMERICAN ASSOCIATION FOR
LABORATORY ACCREDITATION

ACCREDITED LABORATORY

A2LA has accredited

INSPECTION MEASUREMENT COMPANY

Wyoming, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 18 June 2005*).



Presented this 10th day of June 2008.

A handwritten signature in cursive script, reading "Peter Mlynar".

President
For the Accreditation Council
Certificate Number 1030.01
Valid to July 31, 2010

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NSCL Z540-1-1994

INSPECTION MEASUREMENT COMPANY
2291 Byron Center Ave. S.W.
Wyoming, MI 49519-1651
Lester P. Kragt Phone: 616 531 3303

CALIBRATION

Valid To: July 31, 2010

Certificate Number: 1030.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

| Parameter/Equipment | Range | Best Uncertainty ^{2, 6} (\pm) | | Comments |
|-------------------------------------|---|--|--|----------------------------------|
| | | Laboratory | On-Site ³ | |
| Angle Blocks | (0 to 90) ° | 0.30 ° | On-Site calibration service is not available for this parameter. | Optical comparator, angle blocks |
| Bore Gages | (0 to 4.001) in | 0.6R | (0.6R + 10T) μ in | Plain ring gage |
| Calipers ⁴ | (0 to 80) in | 0.6R | (0.6R + 10T) μ in | Plain ring gage and gage blocks |
| Cylindrical Ring Gages ⁴ | (0.04 to 0.3) in (0.3 to 1) in (1 to 10) in | 10 μ in 10 μ in (10 + 9.6L) μ in | On-Site calibration service is not available for this parameter. | Helios LMM; ring masters |
| Depth Micrometers ⁴ | (0 to 12) in | 0.6R | (0.6R + 10T) μ in | Gage blocks |
| Dial Test Indicators ⁴ | To 0.1 in | 0.6R | (0.6R + 10T) μ in | Indicator checker |

| Parameter/Equipment | Range | Best Uncertainty ^{2,4,6} (±) | | Comments |
|--|--|---|--|--------------------------------|
| | | Laboratory | On-Site ³ | |
| Travel Indicators | To 4.0 in | 0. 0.6R | (0.6R + 10T) μin | Indicator checker |
| Feeler Gages | (0 to 0.05) in | 200 μin | On-Site calibration service is not available for this parameter. | Micro Cal |
| Gage Blocks | (0.005 to 1) in (1 to 4) in (5 to 20) in | 4 μin (2.5 + 1.5L) μin (5 + 1.2L) μin | On-Site calibration service is not available for this parameter. | DHC DHC SHC |
| Gage Pins – Class ZZ to Z Class Y to XXX | (0 to 1) in (0 to 10) in | 50 μin (10 + 9.6L) μin | On-Site calibration service is not available for this parameter. | Laser micrometer Helios LMM |
| Spheres | (0 to 5) in | (10 + 9.6D) μin | On-Site calibration service is not available for this parameter. | Helios LMM |
| Granite Surface Plates | To 32 ft ² | (100 + 2.1L) μin | See Footnote 5 | Granite straight edge |
| Snap Gages | (0.30 to 10) in | (10 + 9.6L) μin | On-Site calibration service is not available for this parameter. | Helios LMM, gage blocks |
| Height Gages | (0 to 60) in | (200 + 25L) μin | (10T + 25L + 2000) μin | Gage blocks |
| Inside Micrometers | (0.5 to 20) in | 0.6R | On-Site calibration service is not available for this parameter. | Helios LMM |

| Parameter/Equipment | Range | Best Uncertainty ^{2, 6} (\pm) | | Comments |
|--|--|--|--|---|
| | | Laboratory | On-Site ³ | |
| Micrometer Standards ⁴ | (1 to 20) in (21 to 60) in | (10 + 9.6L) μ in (15 + 1.2L) μ in | On-Site calibration service is not available for this parameter. | Helios LMM SHC |
| Optical Comparators ⁴ & Vision Systems | | | | |
| Magnification | 5X, 10X, 25X, 31.25X, 50X, 62.5X, 100X | (200 + 10T) μ in | See Footnote 5 | Glass scales, Magnification checker gage balls |
| Linear Axis - X & Y | (0 to 6) in (6 to 24) in | (200 + 10T) μ in (34L + 10T) μ in | | |
| Squareness Radius Angularity | (0 to 90) ° | (200 + 10T) μ in (200 + 10T) μ in (0.6R + 10T) | | |
| Outside Micrometers ⁴ | (0 to 20) in | 0.6R | (0.6R + 10T) μ in | Gage blocks, gage ball |
| Protractors | (0 to 90) ° | 0.6R | On-Site calibration service is not available for this parameter. | Optical comparator and angle blocks |
| Radius Gages ⁴ | (¹ / ₁₆ to 1.0) in | 500 μ in | On-Site calibration service is not available for this parameter. | Optical comparator |
| Rules ⁴ | (0 to 96) in | 0.6R | On-Site calibration service is not available for this parameter. | Gage blocks |
| Squares ⁴ – Cylindrical & Solid | (0 to 12) in | 100 μ in | On-Site calibration service is not available for this parameter. | Square master |

| Parameter/Equipment | Range | Best Uncertainty ^{2, 6} (\pm) | | Comments |
|---|---|--|--|---|
| | | Laboratory | On-Site ³ | |
| Thread Plugs ⁴ – Pitch Diameter Major Diameter | (0 to 4) in | 84 μ in 40 μ in | On-Site calibration service is not available for this parameter. | Bench mic with Thread wires Gage blocks |
| Thread Rings ⁴ – Pitch Diameter Minor Diameter | To 2 in dependant on available plugs To 2 in | 300 μ in 300 μ in | On-Site calibration service is not available for this parameter. | Thread set plugs Plain plugs |
| Thread Wires | To 0.3 in | 10 μ in | On-Site calibration service is not available for this parameter. | Helios LMM |
| V-Blocks, Box Parallel, Right Angle | (0 to 12) in | 100 μ in | On-Site calibration service is not available for this parameter. | Square master |
| Linear Digital Scales | (0 to 144) in | $(0.6R + 25L)$ μ in | $(10T + 0.6R + 25L)$ μ in | Gage blocks |
| Tape Measure | (0 to 96) in | $0.6R$ | $(0.6R + 10T)$ μ in | Gage blocks |

II. Mechanical

| Parameter/Equipment | Range | Best Uncertainty ² (±) | | Comments |
|--|---|--|---|---|
| | | Laboratory | On-Site ³ | |
| <p>Direct Verification of Durometers –</p> <p>Verification of indenter shape and extension:</p> <p>Extension at zero reading</p> <p>35 degree circular conical frustum</p> <p>30 degree cone</p> <p>Verification of indenter shape and extension:</p> <p>1.2 mm radius</p> <p>Verification of the durometer spring</p> | <p>Diameter of the base of the frustum</p> <p>Diameter of the top of the frustum</p> <p>Cone angle</p> <p>Diameter of the base of the cone</p> <p>Tip radius</p> <p>Indenter thickness</p> <p>Indenter radius</p> | <p>500 µin</p> <p>500 µin</p> <p>500 µin</p> <p>0.15°</p> <p>500 µin</p> <p>500 µin</p> <p>500 µin</p> <p>500 µin</p> <p>500 µin</p> <p>0.6 grams</p> | <p>On-Site calibration service is not available for this parameter.</p> | <p>Direct verification method per ASTM D2240</p> <p>Verification of these dimensional features is by optical projection.</p> <p>Verification of the spring force is by dead weights. The best uncertainty applies to all durometer types.</p> |
| <p>Indirect Verification of Rockwell & Rockwell Superficial Hardness Testers</p> | <p>HRB (W):</p> <p>Low</p> <p>Middle</p> <p>High</p> <p>HRC:</p> <p>Low</p> <p>Middle</p> <p>High</p> <p>HR15N:</p> <p>Low</p> <p>Middle</p> <p>High</p> | <p>0.6 HRB (W)</p> <p>0.6 HRB (W)</p> <p>1.0 HRB (W)</p> <p>0.6 HRC</p> <p>0.6 HRC</p> <p>0.5 HRC</p> <p>0.6 HR15N</p> <p>0.6 HR15N</p> <p>0.6 HR15N</p> | <p>See Footnote 5</p> | <p>Indirect verification per ASTM E18.</p> |

| Parameter/Equipment | Range | Best Uncertainty ² (±) | | Comments |
|--|--|--|--|-------------------------------------|
| | | Laboratory | On-Site ³ | |
| Indirect Verification of Rockwell & Rockwell Superficial Hardness Testers – (cont) | HR30N: Low Middle High HR15T: Low Middle High HR30T: Low Middle High HR45N: Low Middle High HR45TW: Low Middle High HRA Low Middle High HRF (W) Low Middle High | 0.8 HR30N 0.8 HR30N 1.0 HR30N 1.0 HR15T 0.8 HR15T 0.8 HR15T 1.0 HR30T 0.8 HR30T 0.8 HR30T 1.0 HR45N 0.8 HR45N 0.8 HR45N 1.0 HR45TW 0.8 HR45TW 0.8 HR45TW 0.6 HRA 0.6 HRA 0.6 HRA 0.6 HRF (W) 0.7 HRF (W) 0.6 HRF (W) | See Footnote 5 | Indirect verification per ASTM E18. |
| Scales | (0.5 to 1000) lb | 0.2 % of full scale | See Footnote 5 | Dead weights |
| Torque Wrenches | (10 to 1000) in·lb (20 to 250) ft·lb (60 to 600) ft·lb | 2 % of full scale 2 % of full scale 2 % of full scale | On-Site calibration service is not available for this parameter. | Torque transducer |

| Parameter/Equipment | Range | Best Uncertainty ² (±) | | Comments |
|---------------------|---|-----------------------------------|--|----------------------------------|
| | | Laboratory | On-Site ³ | |
| Force Gages | (0.5 to 500) ft·lb | 2 % of full scale | On-Site calibration service is not available for this parameter. | Class F dead weights |
| Dead Weights | Up to 50 lb | 1 % of reading | On-Site calibration service is not available for this parameter. | Digital balance and dead weights |
| Torque Transducers | (10 to 1000) in·lb (20 to 600) ft·lb | 1 % of full scale | On-Site calibration service is not available for this parameter. | Dead weights |

¹ This laboratory offers commercial and on-site calibration service.

² “Best Uncertainty” is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer’s device, to the environment (if the calibration is performed in the field) and to influences from the circumstances of the specific calibration.

³ On-site calibration service is available for this calibration, where noted. The uncertainties achievable on a customer's site can normally be expected to be larger than the Best Measurement Capabilities (BMC) that the accredited laboratory has been assigned as Best Uncertainty on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the calibration uncertainty being larger than the BMC.

⁴ Metric equivalent devices are available for this parameter.

⁵ The best uncertainty stated for calibrations performed in the laboratory is applicable for calibrations performed on-site.

⁶ In the statement of best uncertainty, L is the numerical value of the nominal length of the device measured in inches; T is the temperature of correction for deviations from 20 °C; R is the numerical value of the resolution of the device in microinches.